

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092624\
 Data File : VU061014.D
 Acq On : 26 Sep 2024 12:22
 Operator : MD/SY
 Sample : VU0926WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK057

Quant Time: Sep 27 04:52:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 27 04:49:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	147132	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	153966	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	57727	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	58035	4.385	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.908	69	51323	5.447	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.000%
11) 1,1-Dichloroethene-d2	2.560	65	20118	4.289	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	4.631	46	116403	54.792	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.580%
24) Chloroform-d	5.055	84	101021	4.994	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.695	65	50100	5.181	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	5.718	84	191584	4.643	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	6.682	67	68202	5.342	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.891	98	149085	3.963	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	8.177	79	20550	4.471	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.627	63	92307	50.915	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.820%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	52966	5.295	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	50621	5.219	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092624\
Data File : VU061014.D
Acq On : 26 Sep 2024 12:22
Operator : MD/SY
Sample : VU0926WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK057

Quant Time: Sep 27 04:52:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 27 04:49:19 2024
Response via : Initial Calibration

