

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092719\
 Data File : VU034851.D
 Acq On : 26 Sep 2019 18:49
 Operator : JC/SP
 Sample : K5007-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CWK30

Quant Time: Sep 27 05:45:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 27 05:10:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	101136	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	94937	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	44331	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	35851	4.46	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.20%
7) Chloroethane-d5	1.67	69	29141	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.20%
11) 1,1-Dichloroethene-d2	2.26	63	54952	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	4.17	46	93761	47.40	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.80%
24) Chloroform-d	4.62	84	53445	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.29	65	30828	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
32) Benzene-d6	5.31	84	111942	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.60%
36) 1,2-Dichloropropane-d6	6.31	67	38879	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.54	98	108024	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	7.83	79	12503	3.62	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.40%
46) 2-Hexanone-d5	8.30	63	67603	44.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.08%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	28176	4.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.20%
64) 1,2-Dichlorobenzene-d4	11.84	152	35632	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds					Ovalue
3) Chloromethane	1.32	50	2728	0.378 ug/L	94
22) cis-1,2-Dichloroethene	4.20	96	16368	2.361 ug/L	97
34) Trichloroethene	6.16	95	2411	0.377 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092719\
Data File : VU034851.D
Acq On : 26 Sep 2019 18:49
Operator : JC/SP
Sample : K5007-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CWK30

Quant Time: Sep 27 05:45:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 27 05:10:18 2019
Response via : Initial Calibration

