

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092719\
 Data File : VU034852.D
 Acq On : 26 Sep 2019 19:14
 Operator : JC/SP
 Sample : K5007-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Sep 27 05:48:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 27 05:10:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	97610	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	92049	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	42676	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	35657	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.67	69	28998	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.00%
11) 1,1-Dichloroethene-d2	2.26	63	51739	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.18	46	91998	48.19	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.38%
24) Chloroform-d	4.62	84	51986	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.29	65	28816	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
32) Benzene-d6	5.32	84	108753	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.31	67	36870	4.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.20%
41) Toluene-d8	7.54	98	101724	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
43) trans-1,3-Dichloropropene-	7.83	79	12680	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.30	63	65605	44.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.16%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	26680	4.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.20%
64) 1,2-Dichlorobenzene-d4	11.84	152	34065	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092719\
Data File : VU034852.D
Acq On : 26 Sep 2019 19:14
Operator : JC/SP
Sample : K5007-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Sep 27 05:48:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 27 05:10:18 2019
Response via : Initial Calibration

