

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044887.D
 Acq On : 27 Sep 2021 13:48
 Operator : SY/MD
 Sample : VU0927WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0927WBL01

Quant Time: Sep 27 16:35:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	197853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	351472	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	323636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	151171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	156679	45.374	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.740%	
35) Dibromofluoromethane	5.295	113	108227	44.886	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.780%	
50) Toluene-d8	7.902	98	444331	48.793	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.580%	
62) 4-Bromofluorobenzene	10.639	95	157941	46.696	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044887.D
Acq On : 27 Sep 2021 13:48
Operator : SY/MD
Sample : VU0927WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0927WBL01

Quant Time: Sep 27 16:35:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

