

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044902.D
 Acq On : 27 Sep 2021 20:26
 Operator : SY/MD
 Sample : M3859-04DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1030-MW-11DL

Quant Time: Sep 28 06:28:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	203141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	360501	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	335220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	161789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	158571	44.726	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.460%	
35) Dibromofluoromethane	5.292	113	108381	43.824	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.640%	
50) Toluene-d8	7.902	98	450959	48.281	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.560%	
62) 4-Bromofluorobenzene	10.635	95	172442	49.707	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.420%	
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	4433	2.217	ug/l	94
27) cis-1,2-Dichloroethene	4.671	96	9880	3.102	ug/l	90
44) Trichloroethene	6.546	130	12104	4.319	ug/l	98
64) Tetrachloroethene	8.555	164	164833	72.544	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044902.D
Acq On : 27 Sep 2021 20:26
Operator : SY/MD
Sample : M3859-04DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1030-MW-11DL

Quant Time: Sep 28 06:28:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

