

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044906.D
 Acq On : 27 Sep 2021 23:10
 Operator : SY/MD
 Sample : VU0927WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0927WBL02

Quant Time: Sep 28 06:29:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	204739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	358680	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.421	117	331743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	154892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	160179	44.827	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.660%	
35) Dibromofluoromethane	5.295	113	110062	44.730	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.460%	
50) Toluene-d8	7.900	98	450648	48.493	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.980%	
62) 4-Bromofluorobenzene	10.636	95	164336	47.610	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044906.D
Acq On : 27 Sep 2021 23:10
Operator : SY/MD
Sample : VU0927WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0927WBL02

Quant Time: Sep 28 06:29:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

