

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044909.D
 Acq On : 28 Sep 2021 00:20
 Operator : SY/MD
 Sample : M3859-01DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1027-MW-01DL

Quant Time: Sep 28 06:30:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	201930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	348857	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	318446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	147973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	154858	43.941	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.880%	
35) Dibromofluoromethane	5.292	113	106822	44.635	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.280%	
50) Toluene-d8	7.903	98	436455	48.288	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.580%	
62) 4-Bromofluorobenzene	10.636	95	164172	48.902	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.800%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.668	96	16707	5.277	ug/l	91
44) Trichloroethene	6.546	130	15540	5.730	ug/l	92
64) Tetrachloroethene	8.555	164	152598	70.696	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044909.D
Acq On : 28 Sep 2021 00:20
Operator : SY/MD
Sample : M3859-01DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1027-MW-01DL

Quant Time: Sep 28 06:30:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

