

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044910.D
 Acq On : 28 Sep 2021 00:44
 Operator : SY/MD
 Sample : M3859-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 1028-MW-02

Quant Time: Sep 28 06:30:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	197689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	348544	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	316990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	148783	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	155149	44.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.94%	
35) Dibromofluoromethane	5.295	113	107495	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
50) Toluene-d8	7.902	98	434046	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	10.635	95	159177	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
Target Compounds					Qvalue	
4) Vinyl Chloride	1.607	62	7390	2.20	ug/l	# 75
16) Acetone	2.636	43	11926	6.13	ug/l	97
21) trans-1,2-Dichloroethene	3.356	96	2748	1.08	ug/l	89
25) 2-Butanone	4.716	43	4564	1.55	ug/l	96
27) cis-1,2-Dichloroethene	4.668	96	103971	33.54	ug/l	87
44) Trichloroethene	6.546	130	125411	46.28	ug/l	98
64) Tetrachloroethene	8.555	164	188364	87.67	ug/l	97
95) Naphthalene	14.092	128	3420	1.24	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044910.D
Acq On : 28 Sep 2021 00:44
Operator : SY/MD
Sample : M3859-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
1028-MW-02

Quant Time: Sep 28 06:30:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

