

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044913.D
 Acq On : 28 Sep 2021 01:54
 Operator : SY/MD
 Sample : M3829-02DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE114D1-20210916DL

Quant Time: Sep 28 06:31:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	192436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	340325	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	314877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	150334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	153482	45.699	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.400%	
35) Dibromofluoromethane	5.292	113	105795	45.314	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.620%	
50) Toluene-d8	7.899	98	425961	48.308	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.620%	
62) 4-Bromofluorobenzene	10.636	95	161536	49.323	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.640%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.575	101	4426	1.891	ug/l	95
27) cis-1,2-Dichloroethene	4.668	96	1239	0.411	ug/l	76
44) Trichloroethene	6.546	130	84399	31.898	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044913.D
Acq On : 28 Sep 2021 01:54
Operator : SY/MD
Sample : M3829-02DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE114D1-20210916DL

Quant Time: Sep 28 06:31:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

