

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044914.D
 Acq On : 28 Sep 2021 02:17
 Operator : SY/MD
 Sample : M3829-04DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE125D1-20210917DL

Quant Time: Sep 28 06:31:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	197769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	348448	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	317083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	148609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	154508	44.764	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.520%	
35) Dibromofluoromethane	5.292	113	108012	45.185	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.380%	
50) Toluene-d8	7.903	98	433780	48.048	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.100%	
62) 4-Bromofluorobenzene	10.636	95	154637	46.116	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.240%	
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.578	101	1777	0.739	ug/l	98
27) cis-1,2-Dichloroethene	4.671	96	986	0.318	ug/l	94
44) Trichloroethene	6.546	130	53613	19.790	ug/l	99
64) Tetrachloroethene	8.558	164	1999	0.930	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044914.D
Acq On : 28 Sep 2021 02:17
Operator : SY/MD
Sample : M3829-04DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE125D1-20210917DL

Quant Time: Sep 28 06:31:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

