

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044916.D
 Acq On : 28 Sep 2021 03:04
 Operator : SY/MD
 Sample : M3829-07DL 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE108D2-20210917DL

Quant Time: Sep 28 06:31:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	198574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	353765	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	319235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	153811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	158945	45.863	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.720%	
35) Dibromofluoromethane	5.292	113	108394	44.664	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.320%	
50) Toluene-d8	7.902	98	439748	47.977	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.960%	
62) 4-Bromofluorobenzene	10.635	95	160295	47.085	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.160%	
Target Compounds						
					Qvalue	
44) Trichloroethene	6.546	130	126053	45.831	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044916.D
Acq On : 28 Sep 2021 03:04
Operator : SY/MD
Sample : M3829-07DL 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE108D2-20210917DL

Quant Time: Sep 28 06:31:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

