

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044918.D
 Acq On : 28 Sep 2021 03:51
 Operator : SY/MD
 Sample : M3829-13DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE120D1-20210917DL

Quant Time: Sep 28 06:31:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	199273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	345835	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	313567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	145241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	155889	44.823	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.640%	
35) Dibromofluoromethane	5.292	113	107363	45.253	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.500%	
50) Toluene-d8	7.900	98	433089	48.334	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.660%	
62) 4-Bromofluorobenzene	10.636	95	159704	47.987	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.980%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	2233	0.921	ug/l	93
12) 1,1-Dichloroethene	2.578	96	1275	0.531	ug/l	88
44) Trichloroethene	6.546	130	162814	60.554	ug/l	98
64) Tetrachloroethene	8.556	164	863	0.406	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044918.D
Acq On : 28 Sep 2021 03:51
Operator : SY/MD
Sample : M3829-13DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE120D1-20210917DL

Quant Time: Sep 28 06:31:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

