

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044922.D
 Acq On : 28 Sep 2021 05:25
 Operator : SY/MD
 Sample : M3915-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE109D1-20210922

Quant Time: Sep 28 06:32:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	193212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	341638	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	318574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	151935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	152585	45.250	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.500%	
35) Dibromofluoromethane	5.292	113	105630	45.070	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.140%	
50) Toluene-d8	7.899	98	426398	48.172	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.340%	
62) 4-Bromofluorobenzene	10.636	95	163873	49.845	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.680%	
Target Compounds						
					Qvalue	
16) Acetone	2.636	43	2110	1.109	ug/l	91
44) Trichloroethene	6.546	130	35724	13.450	ug/l	99
64) Tetrachloroethene	8.555	164	1385	0.641	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044922.D
Acq On : 28 Sep 2021 05:25
Operator : SY/MD
Sample : M3915-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE109D1-20210922

Quant Time: Sep 28 06:32:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

