

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
 Data File : VU044924.D
 Acq On : 28 Sep 2021 06:11
 Operator : SY/MD
 Sample : M3904-02RE
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JET-GROUT-SPOIL-STEEL-WHEEL

Quant Time: Sep 28 08:22:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 25 09:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	190882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	336002	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	316256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	152428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	151585	45.502	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.000%	
35) Dibromofluoromethane	5.292	113	56384	24.461	ug/l	0.00
Spiked Amount	50.000		Recovery	=	48.920%	
50) Toluene-d8	7.899	98	419079	48.139	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.280%	
62) 4-Bromofluorobenzene	10.635	95	165739	51.258	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.520%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	127142	67.660	ug/l	99
18) Methyl Acetate	2.951	43	84317	13.431	ug/l	97
20) Methylene Chloride	3.047	84	9265	3.046	ug/l	93
68) m/p-Xylenes	9.697	106	12603	2.812	ug/l	99
69) o-Xylene	10.102	106	9293	2.144	ug/l	99
95) Naphthalene	14.092	128	3554	1.241	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092721\
Data File : VU044924.D
Acq On : 28 Sep 2021 06:11
Operator : SY/MD
Sample : M3904-02RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JET-GROUT-SPOIL-STEEL-WHEEL

Quant Time: Sep 28 08:22:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092421W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 25 09:30:46 2021
Response via : Initial Calibration

