

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092724\
 Data File : VU061030.D
 Acq On : 27 Sep 2024 13:22
 Operator : MD/SY
 Sample : P4186-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 28 04:05:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 28 04:04:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	132213	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	146229	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	54762	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	51062	4.294	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.800%
7) Chloroethane-d5	1.911	69	45372	5.359	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.200%
11) 1,1-Dichloroethene-d2	2.563	65	18554	4.402	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.000%
20) 2-Butanone-d5	4.634	46	118283	61.959	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.920%
24) Chloroform-d	5.055	84	98501	5.418	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%
26) 1,2-Dichloroethane-d4	5.695	65	48387	5.568	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.400%
32) Benzene-d6	5.721	84	176266	4.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
36) 1,2-Dichloropropane-d6	6.682	67	65492	5.401	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.000%
41) Toluene-d8	7.891	98	134319	3.759	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.200%
43) trans-1,3-Dichloroprop...	8.177	79	19651	4.502	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.627	63	89437	51.942	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.880%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	51578	5.429	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	47282	5.138	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092724\
Data File : VU061030.D
Acq On : 27 Sep 2024 13:22
Operator : MD/SY
Sample : P4186-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 28 04:05:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 28 04:04:44 2024
Response via : Initial Calibration

