

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027246.D
 Acq On : 28 Sep 2018 00:06
 Operator : MD/SY
 Sample : J5029-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-MW-3-8.0-092018

Quant Time: Sep 28 05:00:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	89690	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	152752	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	150295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70059	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	97378	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
35) Dibromofluoromethane	4.88	113	67795	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	7.57	98	249683	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	10.31	95	79542	45.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.88%	
Target Compounds						
40) Benzene	5.39	78	2687	0.434	ug/l	99
69) o-Xylene	9.79	106	1009	0.435	ug/l	90
95) Naphthalene	13.75	128	41362	8.579	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027246.D
Acq On : 28 Sep 2018 00:06
Operator : MD/SY
Sample : J5029-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-MW-3-8.0-092018

Quant Time: Sep 28 05:00:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

