

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027248.D
 Acq On : 28 Sep 2018 00:52
 Operator : MD/SY
 Sample : J5020-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T2-MW-5-8.0-091918

Quant Time: Sep 28 05:10:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	95206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	164778	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	158173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	75623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	100662	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
35) Dibromofluoromethane	4.89	113	69806	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	7.57	98	264831	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	10.31	95	86188	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
Target Compounds						
16) Acetone	2.32	43	26010	9.601	ug/l	99
22) Diisopropyl ether	3.58	45	2083	0.365	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027248.D
Acq On : 28 Sep 2018 00:52
Operator : MD/SY
Sample : J5020-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T2-MW-5-8.0-091918

Quant Time: Sep 28 05:10:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

