

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027290.D
 Acq On : 28 Sep 2018 23:14
 Operator : MD/SY
 Sample : PB113306ZHE#01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB113306ZHE#01

Quant Time: Oct 01 01:38:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	121051	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	202038	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	195376	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	95487	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	111685	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
35) Dibromofluoromethane	4.89	113	81251	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	7.57	98	321251	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
62) 4-Bromofluorobenzene	10.31	95	107814	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
Target Compounds						
20) Methylene Chloride	2.70	84	10510	4.817	ug/l	94
43) Isopropyl Acetate	5.55	43	130474	24.509	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027290.D
Acq On : 28 Sep 2018 23:14
Operator : MD/SY
Sample : PB113306ZHE#01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB113306ZHE#01

Quant Time: Oct 01 01:38:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

