

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027293.D
 Acq On : 29 Sep 2018 00:23
 Operator : MD/SY
 Sample : PB113306ZHE#04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB113306ZHE#04

Quant Time: Oct 01 01:37:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	111362	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	188324	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	182108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	86597	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	105671	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	4.89	113	75267	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
50) Toluene-d8	7.57	98	295229	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
62) 4-Bromofluorobenzene	10.31	95	100602	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
Target Compounds						
20) Methylene Chloride	2.70	84	10070	5.017	ug/l	98
43) Isopropyl Acetate	5.55	43	123549	24.899	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027293.D
Acq On : 29 Sep 2018 00:23
Operator : MD/SY
Sample : PB113306ZHE#04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB113306ZHE#04

Quant Time: Oct 01 01:37:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

