

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027294.D
 Acq On : 29 Sep 2018 00:46
 Operator : MD/SY
 Sample : PB113306ZHE#05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB113306ZHE#05

Quant Time: Sep 29 01:50:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	112573	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	191620	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	182605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	86047	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	107532	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
35) Dibromofluoromethane	4.88	113	76821	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
50) Toluene-d8	7.57	98	302185	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	10.31	95	100270	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
Target Compounds						
20) Methylene Chloride	2.70	84	10113	4.984	ug/l	98
43) Isopropyl Acetate	5.55	43	121588	24.082	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027294.D
Acq On : 29 Sep 2018 00:46
Operator : MD/SY
Sample : PB113306ZHE#05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB113306ZHE#05

Quant Time: Sep 29 01:50:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

