

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027296.D
 Acq On : 29 Sep 2018 01:33
 Operator : MD/SY
 Sample : PB113306ZHE#07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB113306ZHE#07

Quant Time: Oct 01 01:39:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	109741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	187409	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	178806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	84111	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	106475	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	4.88	113	75003	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
50) Toluene-d8	7.57	98	296553	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
62) 4-Bromofluorobenzene	10.31	95	98094	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
Target Compounds						
20) Methylene Chloride	2.70	84	9795	4.952	ug/l	96
43) Isopropyl Acetate	5.55	43	122697	24.848	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027296.D
Acq On : 29 Sep 2018 01:33
Operator : MD/SY
Sample : PB113306ZHE#07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB113306ZHE#07

Quant Time: Oct 01 01:39:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

