

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027301.D
 Acq On : 29 Sep 2018 03:29
 Operator : MD/SY
 Sample : J4776-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EO-01-092718-C

Quant Time: Oct 01 02:26:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	106991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	183102	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	176681	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	85964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	106273	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	4.88	113	75302	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	7.57	98	290641	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
62) 4-Bromofluorobenzene	10.31	95	97888	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.62	43	3529	1.151	ug/l #	83
20) Methylene Chloride	2.70	84	40323	20.909	ug/l	99
30) Chloroform	4.68	83	3427	1.046	ug/l	99
43) Isopropyl Acetate	5.55	43	99868	20.700	ug/l	97
52) Toluene	7.64	92	4907	1.152	ug/l	98
68) m/p-Xylenes	9.38	106	5267	3.164	ug/l	91
84) 1,2,4-Trimethylbenzene	11.15	105	5921	1.078	ug/l	97
95) Naphthalene	13.74	128	271555	38.556	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027301.D
Acq On : 29 Sep 2018 03:29
Operator : MD/SY
Sample : J4776-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EO-01-092718-C

Quant Time: Oct 01 02:26:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

