

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027303.D
 Acq On : 29 Sep 2018 04:15
 Operator : MD/SY
 Sample : J4777-36
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PL-01-092718-B

Quant Time: Oct 01 03:46:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	107478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	184169	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	177652	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	84289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	107100	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
35) Dibromofluoromethane	4.89	113	74909	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	7.57	98	292659	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
62) 4-Bromofluorobenzene	10.31	95	99228	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
Target Compounds						
20) Methylene Chloride	2.70	84	17353	8.957	ug/l	97
43) Isopropyl Acetate	5.55	43	98688	20.337	ug/l	98
68) m/p-Xylenes	9.38	106	823	1.759	ug/l	98
95) Naphthalene	13.75	128	29645	5.794	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027303.D
Acq On : 29 Sep 2018 04:15
Operator : MD/SY
Sample : J4777-36
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PL-01-092718-B

Quant Time: Oct 01 03:46:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

