

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092818\
 Data File : VU027305.D
 Acq On : 29 Sep 2018 05:02
 Operator : MD/SY
 Sample : J4777-40
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PL-01-092718-D

Quant Time: Oct 01 03:52:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	106440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	181528	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	177000	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	85551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	105064	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	4.88	113	74988	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
50) Toluene-d8	7.57	98	289842	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	10.31	95	97430	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
Target Compounds						
18) Methyl Acetate	2.62	43	3182	1.044	ug/l #	89
20) Methylene Chloride	2.70	84	8961	4.671	ug/l	98
43) Isopropyl Acetate	5.55	43	101713	21.265	ug/l	98
95) Naphthalene	13.75	128	7662	2.734	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092818\
Data File : VU027305.D
Acq On : 29 Sep 2018 05:02
Operator : MD/SY
Sample : J4777-40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PL-01-092718-D

Quant Time: Oct 01 03:52:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

