

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044941.D
 Acq On : 28 Sep 2021 15:09
 Operator : SY/MD
 Sample : M3832-01
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-B

Quant Time: Sep 28 15:54:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	215394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	383543	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	364677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	173309	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	170353	47.512	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.292	113	117884	48.722	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	7.899	98	485869	50.824	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.640%
62) 4-Bromofluorobenzene	10.645	95	181777	46.056	ug/l	0.01
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
16) Acetone	2.652	43	2967	1.394	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044941.D
Acq On : 28 Sep 2021 15:09
Operator : SY/MD
Sample : M3832-01
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-B

Quant Time: Sep 28 15:54:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

