

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044945.D
 Acq On : 28 Sep 2021 16:43
 Operator : SY/MD
 Sample : M3953-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SW-RR-03-(SA)

Quant Time: Sep 28 17:19:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	197796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	347689	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	326792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	156194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	155771	47.310	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.620%
35) Dibromofluoromethane	5.295	113	109250	49.810	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	7.902	98	442627	51.075	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.140%
62) 4-Bromofluorobenzene	10.635	95	161915	45.254	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%
Target Compounds						
16) Acetone	2.629	43	3658	1.872	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044945.D
Acq On : 28 Sep 2021 16:43
Operator : SY/MD
Sample : M3953-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SW-RR-03-(SA)

Quant Time: Sep 28 17:19:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

