

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044946.D
 Acq On : 28 Sep 2021 17:07
 Operator : SY/MD
 Sample : M3953-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SW-RR-01-(2.0)

Quant Time: Sep 28 17:45:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	193301	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	342256	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	329958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	161417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	156720	48.705	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.420%
35) Dibromofluoromethane	5.295	113	109809	50.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	7.903	98	436119	51.123	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.240%
62) 4-Bromofluorobenzene	10.636	95	163828	46.516	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.040%
Target Compounds						
16) Acetone	2.633	43	3206	1.679	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044946.D
Acq On : 28 Sep 2021 17:07
Operator : SY/MD
Sample : M3953-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SW-RR-01-(2.0)

Quant Time: Sep 28 17:45:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

