

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044947.D
 Acq On : 28 Sep 2021 17:30
 Operator : SY/MD
 Sample : M3829-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE120D2-20210917

Quant Time: Sep 29 05:46:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	196255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	345701	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	330990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	165522	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	158064	48.38	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.76%
35) Dibromofluoromethane	5.295	113	110249	50.55	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.10%
50) Toluene-d8	7.902	98	441289	51.21	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.42%
62) 4-Bromofluorobenzene	10.635	95	167839	47.18	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.36%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.584	101	42267	17.90	ug/l	99
12) 1,1-Dichloroethene	2.584	96	4669	2.01	ug/l	98
27) cis-1,2-Dichloroethene	4.674	96	7241	2.41	ug/l	86
30) Chloroform	5.099	83	2136	0.42	ug/l	95
38) Carbon Tetrachloride	5.533	117	1625	0.52	ug/l	95
44) Trichloroethene	6.549	130	1308401	515.39	ug/l	98
64) Tetrachloroethene	8.558	164	12822	5.94	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044947.D
Acq On : 28 Sep 2021 17:30
Operator : SY/MD
Sample : M3829-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE120D2-20210917

Quant Time: Sep 29 05:46:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

