

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044949.D
 Acq On : 28 Sep 2021 18:17
 Operator : SY/MD
 Sample : M3962-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 29 05:46:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	192166	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	338095	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	328392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	161989	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	156987	49.076	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.160%
35) Dibromofluoromethane	5.295	113	108527	50.884	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.760%
50) Toluene-d8	7.903	98	434777	51.593	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.180%
62) 4-Bromofluorobenzene	10.636	95	162548	46.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.440%
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	1937	1.020	ug/l	92
44) Trichloroethene	6.549	130	2428	0.978	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044949.D
Acq On : 28 Sep 2021 18:17
Operator : SY/MD
Sample : M3962-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 29 05:46:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

