

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044951.D
 Acq On : 28 Sep 2021 19:05
 Operator : SY/MD
 Sample : M3893-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 29 05:47:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	197809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	347410	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	336995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	166784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	163120	49.539	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.080%
35) Dibromofluoromethane	5.295	113	111060	50.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.360%
50) Toluene-d8	7.903	98	447976	51.734	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.460%
62) 4-Bromofluorobenzene	10.636	95	169499	47.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.820%
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	5401	2.763	ug/l	94
20) Methylene Chloride	3.051	84	2051	0.697	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044951.D
Acq On : 28 Sep 2021 19:05
Operator : SY/MD
Sample : M3893-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 29 05:47:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

