

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044959.D
 Acq On : 28 Sep 2021 23:00
 Operator : SY/MD
 Sample : VU0928WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0928WBL02

Quant Time: Sep 29 05:48:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	192705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	342182	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	327980	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	162181	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	157663	49.150	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.300%
35) Dibromofluoromethane	5.295	113	109149	50.565	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.120%
50) Toluene-d8	7.903	98	438913	51.461	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.920%
62) 4-Bromofluorobenzene	10.636	95	162132	46.044	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044959.D
Acq On : 28 Sep 2021 23:00
Operator : SY/MD
Sample : VU0928WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0928WBL02

Quant Time: Sep 29 05:48:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

