

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044965.D
 Acq On : 29 Sep 2021 01:21
 Operator : SY/MD
 Sample : M3451-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-20210924

Quant Time: Sep 29 05:50:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	190212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	337939	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	333917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	160904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	156403	49.396	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.800%
35) Dibromofluoromethane	5.295	113	108716	50.997	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	7.902	98	440939	52.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.700%
62) 4-Bromofluorobenzene	10.636	95	164942	47.430	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.860%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	65497	34.849	ug/l	97
20) Methylene Chloride	3.054	84	17597	6.216	ug/l	95
25) 2-Butanone	4.710	43	13986	5.238	ug/l	96
43) Isopropyl Acetate	5.915	43	21123	3.003	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044965.D
Acq On : 29 Sep 2021 01:21
Operator : SY/MD
Sample : M3451-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-20210924

Quant Time: Sep 29 05:50:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

