

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044968.D
 Acq On : 29 Sep 2021 02:32
 Operator : SY/MD
 Sample : M3920-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-092321

Quant Time: Sep 29 05:50:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	188451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	334891	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	331041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	165836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	154524	49.258	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.520%
35) Dibromofluoromethane	5.295	113	108125	51.181	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	7.902	98	435248	52.143	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.280%
62) 4-Bromofluorobenzene	10.635	95	167009	48.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	2812	1.510	ug/l	96
78) n-propylbenzene	10.912	91	5355	0.356	ug/l	97
84) 1,2,4-Trimethylbenzene	11.475	105	6487	0.606	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044968.D
Acq On : 29 Sep 2021 02:32
Operator : SY/MD
Sample : M3920-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-092321

Quant Time: Sep 29 05:50:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

