

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044972.D
 Acq On : 29 Sep 2021 04:06
 Operator : SY/MD
 Sample : M3936-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 P-1

Quant Time: Sep 29 05:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	191462	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	335719	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	328390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	159745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	157186	49.319	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.640%
35) Dibromofluoromethane	5.295	113	110704	52.272	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.540%
50) Toluene-d8	7.902	98	436960	52.219	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.440%
62) 4-Bromofluorobenzene	10.635	95	163023	47.189	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.380%
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	2059	1.088	ug/l	94
30) Chloroform	5.096	83	4940	0.995	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044972.D
Acq On : 29 Sep 2021 04:06
Operator : SY/MD
Sample : M3936-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
P-1

Quant Time: Sep 29 05:51:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

