

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044975.D
 Acq On : 29 Sep 2021 05:17
 Operator : SY/MD
 Sample : M3952-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1D

Quant Time: Sep 29 07:58:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	192353	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	342447	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	338427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	168878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	160010	49.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.940%
35) Dibromofluoromethane	5.295	113	108587	50.265	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.540%
50) Toluene-d8	7.903	98	444267	52.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.100%
62) 4-Bromofluorobenzene	10.636	95	168787	47.897	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.800%
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	2468	1.299	ug/l	97
39) Methylcyclohexane	6.771	83	1052	0.239	ug/l	85
64) Tetrachloroethene	8.558	164	3986	1.805	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044975.D
Acq On : 29 Sep 2021 05:17
Operator : SY/MD
Sample : M3952-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1D

Quant Time: Sep 29 07:58:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

