

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044976.D
 Acq On : 29 Sep 2021 05:40
 Operator : SY/MD
 Sample : M3952-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2

Quant Time: Sep 29 07:58:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	188439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	339115	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	331502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	163482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	158305	50.467	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.940%	
35) Dibromofluoromethane	5.295	113	109138	51.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.040%	
50) Toluene-d8	7.902	98	435123	51.478	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.960%	
62) 4-Bromofluorobenzene	10.635	95	165128	47.319	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.640%	
Target Compounds						
					Qvalue	
16) Acetone	2.636	43	2752	1.478	ug/l	99
64) Tetrachloroethene	8.558	164	624	0.288	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044976.D
Acq On : 29 Sep 2021 05:40
Operator : SY/MD
Sample : M3952-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2

Quant Time: Sep 29 07:58:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

