

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044981.D
 Acq On : 29 Sep 2021 07:38
 Operator : SY/MD
 Sample : M3926-04DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-1-MW-4-0921DL

Quant Time: Sep 29 08:00:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	185848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	332116	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	328563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	165040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	154888	50.066	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.140%			
35) Dibromofluoromethane	5.295	113	108643	51.856	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.720%			
50) Toluene-d8	7.903	98	432574	52.255	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.520%			
62) 4-Bromofluorobenzene	10.636	95	164208	48.047	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.100%			
Target Compounds						Qvalue
67) Ethyl Benzene	9.575	91	10912	0.861	ug/l	91
68) m/p-Xylenes	9.700	106	3203	0.650	ug/l	96
73) Isopropylbenzene	10.488	105	35979	2.927	ug/l	99
78) n-propylbenzene	10.909	91	70182	4.686	ug/l	99
80) 1,3,5-Trimethylbenzene	11.092	105	8858	0.823	ug/l	94
83) tert-Butylbenzene	11.423	119	4863	0.464	ug/l	87
84) 1,2,4-Trimethylbenzene	11.472	105	364492	34.219	ug/l	100
85) sec-Butylbenzene	11.648	105	22835	1.693	ug/l	99
89) n-Butylbenzene	12.211	91	6618	0.662	ug/l #	71
95) Naphthalene	14.092	128	18070	1.484	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
Data File : VU044981.D
Acq On : 29 Sep 2021 07:38
Operator : SY/MD
Sample : M3926-04DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-1-MW-4-0921DL

Quant Time: Sep 29 08:00:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

