



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092821\
 Data File : VU044957.D
 Acq On : 28 Sep 2021 21:49
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Title : SW846 8260
 Last Update : Tue Sep 28 12:32:13 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	8261	PASS
75	95	30	60	53.5	20997	PASS
95	95	100	100	100.0	39237	PASS
96	95	5	9	6.7	2638	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	70.6	27693	PASS
175	174	5	9	7.5	2072	PASS
176	174	95	101	96.0	26581	PASS
177	176	5	9	6.6	1743	PASS