

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044987.D
 Acq On : 29 Sep 2021 12:45
 Operator : SY/MD
 Sample : VU0929WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0929WBL01

Quant Time: Sep 29 14:21:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	190656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	342223	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	333402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	165385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	158822	50.04	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.08%
35) Dibromofluoromethane	5.295	113	110486	51.18	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.36%
50) Toluene-d8	7.902	98	444026	52.05	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.10%
62) 4-Bromofluorobenzene	10.635	95	167030	47.43	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.86%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044987.D
Acq On : 29 Sep 2021 12:45
Operator : SY/MD
Sample : VU0929WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0929WBL01

Quant Time: Sep 29 14:21:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

