

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044991.D
 Acq On : 29 Sep 2021 14:30
 Operator : SY/MD
 Sample : M3962-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 29 14:45:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	191025	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	344528	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	336633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	166211	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	159236	50.08	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.16%	
35) Dibromofluoromethane	5.295	113	110384	50.79	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.58%	
50) Toluene-d8	7.903	98	444695	51.78	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.56%	
62) 4-Bromofluorobenzene	10.636	95	166218	46.88	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044991.D
Acq On : 29 Sep 2021 14:30
Operator : SY/MD
Sample : M3962-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 29 14:45:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

