

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044993.D
 Acq On : 29 Sep 2021 15:17
 Operator : SY/MD
 Sample : M3870-01
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-C

Quant Time: Sep 29 16:08:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	204092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	365768	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	369180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	186400	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	166159	48.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.82%
35) Dibromofluoromethane	5.292	113	112855	48.91	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.82%
50) Toluene-d8	7.899	98	465836	51.10	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.20%
62) 4-Bromofluorobenzene	10.639	95	192594	51.17	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.34%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044993.D
Acq On : 29 Sep 2021 15:17
Operator : SY/MD
Sample : M3870-01
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-C

Quant Time: Sep 29 16:08:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

