

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044994.D
 Acq On : 29 Sep 2021 15:40
 Operator : SY/MD
 Sample : M3872-01ME
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SP-LPA-01ME

Manual Integrations
 APPROVED

MMDadoda
 9/30/2021 12:01:29 PM

Quant Time: Sep 29 16:08:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	204950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	371043	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	375837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	191841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	168263	49.32	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.64%		
35) Dibromofluoromethane	5.292	113	113395	48.45	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.90%		
50) Toluene-d8	7.899	98	478867	51.78	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	103.56%		
62) 4-Bromofluorobenzene	10.639	95	200964	52.63	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.26%		
Target Compounds						Qvalue
39) Methylcyclohexane	6.761	83	1046m	0.22	ug/l	
67) Ethyl Benzene	9.574	91	5007	0.35	ug/l	98
68) m/p-Xylenes	9.703	106	1721	0.31	ug/l	81
73) Isopropylbenzene	10.491	105	3749	0.26	ug/l	99
78) n-propylbenzene	10.912	91	4178	0.24	ug/l	99
80) 1,3,5-Trimethylbenzene	11.095	105	16310	1.30	ug/l	96
84) 1,2,4-Trimethylbenzene	11.475	105	37860	3.06	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044994.D
Acq On : 29 Sep 2021 15:40
Operator : SY/MD
Sample : M3872-01ME
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SP-LPA-01ME

Manual Integrations
APPROVED

MMDadoda
9/30/2021 12:01:29 PM

Quant Time: Sep 29 16:08:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

