

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044995.D
 Acq On : 29 Sep 2021 16:04
 Operator : SY/MD
 Sample : M3889-01ME
 Misc : 6.50g/5mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SUP-UST-03-(5-7)ME

Quant Time: Sep 29 16:39:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	212368	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	390133	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	399698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	207589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	174462	49.35	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 98.70%			
35) Dibromofluoromethane	5.292	113	117695	47.82	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.64%			
50) Toluene-d8	7.902	98	512687	52.72	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.44%			
62) 4-Bromofluorobenzene	10.639	95	223957	55.78	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 111.56%			
Target Compounds						
						Qvalue
39) Methylcyclohexane	6.761	83	61541	12.29	ug/l	97
73) Isopropylbenzene	10.491	105	81541	5.27	ug/l	100
78) n-propylbenzene	10.912	91	185684	9.86	ug/l	99
85) sec-Butylbenzene	11.648	105	202932	11.96	ug/l	100
89) n-Butylbenzene	12.214	91	269799	21.47	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044995.D
Acq On : 29 Sep 2021 16:04
Operator : SY/MD
Sample : M3889-01ME
Misc : 6.50g/5mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SUP-UST-03-(5-7)ME

Quant Time: Sep 29 16:39:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

