

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044996.D
 Acq On : 29 Sep 2021 16:28
 Operator : SY/MD
 Sample : M3889-02ME
 Misc : 7.04g/5mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SUP-UST-03-(8-10)ME

Quant Time: Sep 29 17:32:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	225136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	409612	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	415360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	216223	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	179777	47.97	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.94%		
35) Dibromofluoromethane	5.292	113	122321	47.34	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.68%		
50) Toluene-d8	7.899	98	535582	52.46	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	104.92%		
62) 4-Bromofluorobenzene	10.639	95	229911	54.54	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.08%		
Target Compounds						
31) Cyclohexane	5.375	56	32471	5.70	ug/l #	81
39) Methylcyclohexane	6.761	83	86137	16.38	ug/l	97
73) Isopropylbenzene	10.491	105	39910	2.48	ug/l	100
78) n-propylbenzene	10.912	91	74520	3.80	ug/l	98
85) sec-Butylbenzene	11.648	105	96008	5.43	ug/l	99
89) n-Butylbenzene	12.211	91	120280	9.19	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044996.D
Acq On : 29 Sep 2021 16:28
Operator : SY/MD
Sample : M3889-02ME
Misc : 7.04g/5mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SUP-UST-03-(8-10)ME

Quant Time: Sep 29 17:32:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

