

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU044998.D
 Acq On : 29 Sep 2021 17:15
 Operator : SY/MD
 Sample : M3936-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 P-1

Quant Time: Sep 30 04:32:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	219803	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	393714	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	376845	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	183054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	173155	47.32	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.64%
35) Dibromofluoromethane	5.292	113	120990	48.71	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.42%
50) Toluene-d8	7.903	98	506414	51.60	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.20%
62) 4-Bromofluorobenzene	10.636	95	189891	46.87	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.74%
Target Compounds						
30) Chloroform	5.092	83	5754	1.01	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU044998.D
Acq On : 29 Sep 2021 17:15
Operator : SY/MD
Sample : M3936-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
P-1

Quant Time: Sep 30 04:32:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

