

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045000.D
 Acq On : 29 Sep 2021 18:02
 Operator : SY/MD
 Sample : M3936-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-9-24-21

Quant Time: Sep 30 04:33:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	212300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	383209	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	357112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	166106	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	168335	47.63	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.26%
35) Dibromofluoromethane	5.295	113	117520	48.61	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.22%
50) Toluene-d8	7.902	98	484575	50.73	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.46%
62) 4-Bromofluorobenzene	10.635	95	175346	44.47	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.94%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045000.D
Acq On : 29 Sep 2021 18:02
Operator : SY/MD
Sample : M3936-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-9-24-21

Quant Time: Sep 30 04:33:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

