

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045004.D
 Acq On : 29 Sep 2021 19:36
 Operator : SY/MD
 Sample : M3920-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-092321

Quant Time: Sep 30 04:33:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	205649	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	363226	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	331203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	140767	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	161477	47.17	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.34%
35) Dibromofluoromethane	5.295	113	113534	49.55	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.10%
50) Toluene-d8	7.903	98	461721	51.00	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.00%
62) 4-Bromofluorobenzene	10.636	95	157367	42.10	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.20%
Target Compounds						
11) Tert butyl alcohol	3.189	59	28941	29.46	ug/l	# 76
16) Acetone	2.630	43	17359	8.54	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045004.D
Acq On : 29 Sep 2021 19:36
Operator : SY/MD
Sample : M3920-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-092321

Quant Time: Sep 30 04:33:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

