

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045005.D
 Acq On : 29 Sep 2021 20:00
 Operator : SY/MD
 Sample : M3926-02DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-1-MW-2-0921DL

Quant Time: Sep 30 04:33:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	204716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	364347	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	342944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	162382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	161758	47.47	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.94%		
35) Dibromofluoromethane	5.295	113	112617	49.00	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.00%		
50) Toluene-d8	7.903	98	460298	50.69	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.38%		
62) 4-Bromofluorobenzene	10.636	95	168756	45.01	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.02%		
Target Compounds						
						Qvalue
67) Ethyl Benzene	9.575	91	66220	5.00	ug/l	100
68) m/p-Xylenes	9.700	106	22541	4.39	ug/l	100
73) Isopropylbenzene	10.488	105	102924	8.51	ug/l	100
78) n-propylbenzene	10.909	91	203660	13.82	ug/l	100
80) 1,3,5-Trimethylbenzene	11.092	105	47669	4.50	ug/l	99
83) tert-Butylbenzene	11.423	119	8476	0.82	ug/l	87
84) 1,2,4-Trimethylbenzene	11.472	105	828482	79.05	ug/l	99
85) sec-Butylbenzene	11.648	105	44038	3.32	ug/l	98
86) p-Isopropyltoluene	11.796	119	10760	1.01	ug/l	94
89) n-Butylbenzene	12.211	91	23366	2.38	ug/l #	69
95) Naphthalene	14.089	128	36399	3.04	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045005.D
Acq On : 29 Sep 2021 20:00
Operator : SY/MD
Sample : M3926-02DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-1-MW-2-0921DL

Quant Time: Sep 30 04:33:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

