

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045007.D
 Acq On : 29 Sep 2021 20:47
 Operator : SY/MD
 Sample : M3926-04DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-1-MW-4-0921DL

Quant Time: Sep 30 04:34:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	203691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	366372	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	342248	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	160973	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	162730	47.99	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.98%		
35) Dibromofluoromethane	5.295	113	113046	48.91	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.82%		
50) Toluene-d8	7.903	98	463215	50.72	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.44%		
62) 4-Bromofluorobenzene	10.636	95	170692	45.27	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.54%		
Target Compounds						Qvalue
67) Ethyl Benzene	9.575	91	20595	1.56	ug/l	100
68) m/p-Xylenes	9.697	106	5835	1.14	ug/l	93
69) o-Xylene	10.105	106	1275	0.25	ug/l	83
73) Isopropylbenzene	10.488	105	74880	6.25	ug/l	100
78) n-propylbenzene	10.909	91	141633	9.70	ug/l	100
80) 1,3,5-Trimethylbenzene	11.092	105	19229	1.83	ug/l	100
83) tert-Butylbenzene	11.423	119	10583	1.04	ug/l	92
84) 1,2,4-Trimethylbenzene	11.472	105	757852	72.94	ug/l	99
85) sec-Butylbenzene	11.645	105	46483	3.53	ug/l	98
89) n-Butylbenzene	12.214	91	12421	1.27	ug/l #	68
95) Naphthalene	14.089	128	48763	4.11	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045007.D
Acq On : 29 Sep 2021 20:47
Operator : SY/MD
Sample : M3926-04DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-1-MW-4-0921DL

Quant Time: Sep 30 04:34:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

