

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
 Data File : VU045008.D
 Acq On : 29 Sep 2021 21:10
 Operator : SY/MD
 Sample : M3926-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-1-MW-1-0921

Quant Time: Sep 30 04:34:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 12:32:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	206529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	369036	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	341398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	160255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	162797	47.353	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.700%		
35) Dibromofluoromethane	5.292	113	113179	48.616	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.240%		
50) Toluene-d8	7.903	98	464270	50.473	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.940%		
62) 4-Bromofluorobenzene	10.636	95	175763	46.283	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.560%		
Target Compounds						
					Qvalue	
16) Acetone	2.630	43	3896	1.909	ug/l	97
17) Carbon Disulfide	2.797	76	4198	0.581	ug/l	96
39) Methylcyclohexane	6.764	83	2899	0.612	ug/l	98
67) Ethyl Benzene	9.571	91	60276	4.576	ug/l	97
68) m/p-Xylenes	9.700	106	19068	3.726	ug/l	98
69) o-Xylene	10.105	106	4060	0.799	ug/l	71
73) Isopropylbenzene	10.488	105	2031281	170.209	ug/l	100
78) n-propylbenzene	10.909	91	4382104	301.318	ug/l	100
83) tert-Butylbenzene	11.423	119	197271	19.399	ug/l	89
84) 1,2,4-Trimethylbenzene	11.472	105	5033396	486.646	ug/l	99
85) sec-Butylbenzene	11.648	105	918326	70.125	ug/l	99
89) n-Butylbenzene	12.211	91	378802	39.042	ug/l #	79
95) Naphthalene	14.089	128	119929	10.143	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092921\
Data File : VU045008.D
Acq On : 29 Sep 2021 21:10
Operator : SY/MD
Sample : M3926-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-1-MW-1-0921

Quant Time: Sep 30 04:34:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U092821W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 12:32:13 2021
Response via : Initial Calibration

